
REPORTING 6.4.5

Report pH measurements in the data base to the nearest 0.1 standard pH unit.

- ▶ pH values to the nearest 0.05 unit can be published provided the instrument system has been calibrated at the required precision and accuracy.
- ▶ Report the instrument accuracy range with the published values and record the accuracy range in the data base, if possible.
- ▶ Enter the field-determined pH under the correct parameter code on the NWQL analytical services request form and on field forms.